Ш



CERTIFICATE

No. Z1 17 06 01310 001

Holder of Certificate: Rehabilitations-Systeme AG

> Roggenstrasse 3 4665 Oftringen **SWITZERLAND**

Production Facility(ies):

94700

Certification Mark:



Product:

Scooters

Model(s):

Gatsby

Parameters:

Class of use:

1565 mm Total lenght: Total width: 670 mm

1075 mm - 1120 mm Total hight:

Obstacle height: 100 mm 15 km/h Max. speed: 150 kg Max. load:

8° (in the standard seat position Max. safe slope:

as specified by the manufacturer)

Driving motor: 24 VDC, 950 W, 4200 RPM

C

Power electronics: PG Drives Technology, type: S-drive

(D51445.06)

2x (12 VDC, 55 Ah (C20)) Batteries: IN: 100-240 VAC, 50/60 Hz, 4 A;

Battery charger:

OUT: 24 VD

Tested according to: DIN EN 12184:2014

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

713090519

Valid until:

2022-07-02

Date, 2017-07-03

(Torsten Zimmer)

Page 1 of 1

