

CE



USER MANUAL

Our passion – your independence!



CAUTION! Manufacturer takes responsibility for proper usage of the device only when the product was purchased from qualified Akces-Med employee or at specialized medical store!

CAUTION! AKCES-MED Ltd. reserve the right to introduce technical and trading changes in the content of the instruction without warning.

Medical product first class according to the Regulation 93/42/EWG of European Parliament and Council regarding medical products. For this product the producer issued relevant declaration of conformity.

Akces-Med. Sp. z o.o. Company introduced quality management system ISO 13485 regarding production, sales and service. The system is certified by Det Norske Veritas.







CONTENTS

1. The purpose of user manual	6
2. Indications for use	6
3. General characteristics of stroller	7
3.1. Standard equipment	7
3.2. Additional equipment	8
4. Size table	12
5. Detailed description how to operate stroller	13
5.1 Floding the stroller	13
5.2 Unfloding the stroller	14
5.3 Footrest	15
5.4 Wheels	16
5.5 The push handle	17
5.6 Backrest	17
5.7 Przedłużenie oparcia	18
5.8 Seat	18
5.9 5-points belts	19
6. Additional equipment	21
6.1 Safety rail	21
6.2 Thigh abduction belt	21
6.3 Trunk stabilizing belt	23
6.4 Calf belt	25
6.5 Head-neck support	25
6.6 Loins support	26
6.7 Trunk belt	26
6.8 Abduction block	27
6.9 Thigh belt	28
6.10 Narrowing pads (set) 6 cm and 10 cm	28





6.11 High rigid trunk support	29
6.12 Low latteral support	
Low latteral support with seat narrowing cushion 6 cm	
Low latteral support with seat narrowing cushion 10 cm	30
6.13 H vest, 4-points safet vest	31
6.14 6-points safety vest	33
6.15 Umbrella	35
6.16 Stolik	35
6.17 Canopy	36
6.18 Cup holder	37
6.19 Elastico cushion seat	
Seat cushion (thighs shape)	
Seat cushion (pommel shape)	38
6.20 Tilt angle adjustment of the footrest	38
6.21 Bags	39
6.22 Under seat basket	40
7. The use of stroller	41
7.1. Preparation of the patient	41
7.2. How to use stroller	42
8. Safety rules	43
9. Maintenance and cleaning	46





INTRODUCTION

ULISES EVO™ is a perfect stroller designed for daily rehabilitation of children and youth with locomotor dysfunctions. It is particularly useful for patients requiring additional stabilization. The stroller has adjustable seat depth and backrest tilt angle which greatly facilitates its functionality.

With simple and light construction it can be easily folded for transporting purposes. Furthermore, relatively small size allows you to fit **ULISES EVO™** into an undersized car trunk. Stable structure of the stroller, belts and supports system make the user feel safe. The backrest can be easily bend back and adjustable push handle increases the stroller maneuverability. Rich equipment and wide range of adjustments make **ULISES EVO™** fit easily to the individual needs.

If you have any questions or remarks, do not hesitate to contact with us: **Medical consultant:** mobile +48 508 382 509 **Akces-MED Sp. z o.o.:** ph. +48 17 864 04 77 **e-mail:** export@akces-med.com

Thank you for purchasing our product.is a solid device which makes it possible for patient to adopt



CAUTION! Before using the product, user is obliged to familiarize himself/herself with user manual. Remember that complying with guidelines is highly important. It assures safety and extends durability as well as aesthetics of the device.





1 THE PURPOSE OF USER

The present user manual contains basic information which is essential for proper usage, preparation, maintenance, cleaning and guarantee conditions regarding **ULISES EVO™** stroller.

User manual should be kept in easily accessible place. It is designed for people who look after disabled children, and for both doctors as well as physiotherapists who operate the device.



CAUTION! Before using the product, user is obliged to familiarize himself/herself with user manual.

Remember that complying with guidelines is highly important. It assures safety and extends durability as well as aesthetics of the device.

2 INDICATIONS FOR USE

The device is recommended for people in the following cases:

- Cerebral palsy various types and forms.
- Brain and spinal cord injuries with lower limbs paralysis and weakness, balance and motor coordination disorders with different etiology.
- Demyelinating disease (e.g. multiple sclerosis SM).
- Meningocele.
- Muscles diseases with weakness and paralysis (dystrophy, myasthenia gravis, myositis).
- Genetic disorders.
- Degeneration and neurological diseases.
- Other condition connected with muscles weakness, paralysis and musculoskeletal disorders.
- After head and spinal cord injuries and spine trauma with paralysis and paresis.





3 GENERAL CHARACTERISTICS OF STROLLER

3.1 Standard equipment

- 1. Adjustable push handle.
- 2. Headrest.
- 3. 5-point belts.
- 4. Side covers.
- 5. Backrest with tilt angle adjustment.
- 6. Upholstery with the system regulating the tension of seat and backrest upholstery.
- 7. Seat with depth adjustment.
- 8. Footrest with height and tilt angle adjustments, 2-point belts fastening the feet.
- 9. Foldable stroller frame.
- 10. Rear, air pumped wheels equipped with two independent brakes.
- 11. Front swivel castors
- 12. Backrest extension







3.2. Additional equipment















🔜 www.akces-med.com 🛛 🕾 +48 17 864 04 77 🛛 🖂 export@akces-med.com













SIZE TABLE



Symbol	Measurement	Unit	Size 1	Size 1a	Size 2	Size 2a	Size 3
А	Seat width	[cm]	31	31	37	37	43
В	Seat depth	[cm]	22-29	29-37	29-37	35-45	35-45
с	Backrest height (without backrest extension)	[cm]	46	54	54	62	62
C1	Backrest height (with backrest extension)	[cm]	61-70	69-80	69-80	77-90	77-90
D	Backrest headrest	[cm]	40-70	43-80	43-80	51-90	51-90
E	Backrest width	[cm]	28	28	34	34	40
F	Footrests length	[cm]	19-25	24-32	24-32	30-40	30-40
	Footplate length	[cm]	19	19	21	21	25
	Push handle height	[cm]	102	102	102	102	102
х	Backrest tilt angle	[°]	90-150	90-150	90-150	90-150	90-150
Y	Seat tilt angle	[°]	8-18	8-18	8-18	8-18	8-18
z	Footrests tilt angle	[°]	90-180	90-180	90-180	90-180	90-180
	Maximum user weight	[kg]	55	75	75	75	75

Weight and dimensions of the device		Size 1	Size 1a	Size 2	Size 2a	Size 3
Width	[cm]	57	57	63	63	69
After folding width	[cm]	57	57	63	63	69
Length	[cm]	94	110	110	127	127
After folding length	[cm]	104	107	107	114	114
Height	[cm]	102	102	102	102	102
After folding height	[cm]	40	45	45	50	50
Weight	[kg]	18,5	19,5	20	21,5	22

Note: Dimensional tolerance in the table is +1/-1 cm.







DETAILED DESCRIPTION HOW TO USE STROLLER

5.1. Folding the stroller

To fold the stroller, you should:

STEP 1. Fold the footplate (fig. 1, 2).



STEP 2. Change the position of the footplate so that, to be placed in one line with the stroller seat (fig.3) (Footrest **angle adjustment** → see page 16, point 5.3.2).

STEP 3. Pull the strap located in the back of the backrest (fig.4) and simultaneously push the backrest forward until it stops (fig.5) (Backrest tilt angle adjustment → see page 17, point 5.6.1).













STEP 4. In order to make stroller folding easier, it is recommended to take off the side covers.



STEP 5. Afterwards, pull the folding mechanism (fig.7) and push the backrest forward until it stops (fig.8-10).





5.2. Unfolding the stroller

To unfold the stroller, you should:

STEP 1. Pull the stroller's handle backwards until the folding mechanism snaps.









CAUTION! Make sure whether the stroller has been properly unfolded and the folding mechanism is totally locked.

STEP 2. Afterwards, set the desired angle of backrest and footrest, unfold the footrest, put the side covers.

5.3. Footrest

5.3.1 To adjust the footrest height, you should:

STEP 1. Remove the quick connectors from both sides of the stroller. (fig.13);

STEP 2. Set the desired height of the footrest (fig.14);

STEP 3. Put the quick connectors into holes;









5.3.2. In order to adjust the footrest tilt angle, you should simultaneously press the buttons placed on the side of the footrest, set the desired position, and then release the buttons.





5.3.2. The footrest comes standard with 2-point belts stabilizing the feet.

5.4. Wheels

5.4.1. The stroller is equipped with front swivel castors (fig.17).

5.4.2. Air pumped rear wheels.

5.4.3. The stroller has an independent lock for each rear wheel. To activate brake, pull the red lever located at the wheel backwards until it stops (fig.18). This action should be repeated with the other rear wheel.

5.4.4. The brakes have adjustable breaking force. In order to adjust it, you should tighten or loosen the hex nut located at the brake.









5.5. The push handle

5.5.1. In order to adjust the push handle tilt angle, press the two buttons located at the outer side of the push handle simultaneously (fig.19) and then set the desired angle of the push handle.

5.6. Backrest

5.6.1. In order to adjust the backrest tilt angle, you should:

STEP 1. Pull the belt located at the back of the backrest (fig.20) and then set the desired angle of the backrest (fig.21).

STEP 2. Release the belt and make sure that the backrest is totally locked.

Adjustment of the backrest tilt angle is gradually. Before the use of the stroller make sure that the retaining pin is placed in hole and the backrest is locked (fig.22).

5.7. Backrest extension

5.7.1. In order to adjust the backrest extension height, you should:

STEP 1. Using the Allen wrench, loosen the screws located at the back of the backrest (fig.23).

STEP 2. Set the desired height of the backrest extension. **STEP 3.** Tighten the screws.















5.8 Seat

5.8.1. In order to adjust the seat depth, you should:

STEP 1. Using the Allen wrench, loosen the holder located at the both side of the seat frame (fig.25).

STEP 2. Detach upholstery Velcro straps located under the seat and behind the backrest.

STEP 3. Set the desired seat depth.

STEP 4. Using Allen wrench, tighten the holder and fasten the Velcro straps.





Maximum depth setting of the seat is indicated by the retaining pins coming into holes located at the both side of the seat frame. To decrease the seat depth, after reaching its maximum position, push the retaining pins inside the frame.

5.8.2 In order to adjust the angle of the seat, there is necessary to:

- **STEP 1.** Detach Velcros on the straps located below the seat (fig.27).
- STEP 2. Adjust required angle by loosing or tensing the upholster.

STEP 3. Attach the Velcros (fig.28).









5.9. 5-points belts

The stroller is equipped with 5-point belts that stabilize the patient and secure him from falling out of the stroller. Depending on the needs of the patient, the belts can be attached at the different height of the backrest.

STEP 1. In order to attach the belts to the lower part of the backrest, interlace the belts through holes placed in the stroller frame (fig.27) and then insert their ends into strap buckle (fig.28).





STEP 2. In order to attach the belts to the higher part of the backrest, insert ends of the belts into holes placed on the backrest. Afterwards, secure them using knob placed on the other side.

STEP 3. Connect the belts (fig.29) and insert them into buckle located at the seat (fig.30).







STEP 4. In order to attach the 5-point belts to the seat, interlace the belt through the plastic holder located at the cushion

Properly attached belts were shown in the picture (fig.31).



CAUTION! The belt should be interlaced through the plastic holder three times! Assembly diagram below.

ASSEMBLY DIAGRAM OF THE BELTS











6 ADDITIONAL EQUIPMENT

6.1. REF ULE_013 Safety rail

In order to mount the safety rail, you should:

STEP 1. Place the safety rail hooks on the stroller frame, as indicated in the picture 32.



STEP 2. Set the safety rail at the desired position. To do that, you need to release the latch so that the pin jumped into the frame hole.





6.2. **REF ULE_101** Thigh abduction belt

In order to mount the Thigh abduction belt, you should:

STEP 1. Detach the two plastic strap adjusters (pic.35).

STEP 2. Put the belts on the seat (breathable fabric should be facing up, pic. 36).











STEP 3. Attach the belt to the black holder placed next to the backrest (pic. 37).

STEP 4. Slide the end of the belt into the plastic strap adjuster (pic. 38-39).



STEP 5. Then, attach the plastic strap adjuster (the one that you previously detached in STEP 1) to the belt (pic. 40).







STEP 6. Slide the end of the belt into the snap buckle (pic. 41-42).







🛿 www.akces-med.com 🛛 🕾 +48 17 864 04 70 🛛 produkcja@akces-med.com





STEP 7. Finally, slide again the end of the belt into the plastic strap adjusters





6.3. REF ULE_113 Trunk stabilizing belt

In order to mount the trunk stabilizing belt, you should:

STEP 1. Put the straps into the holes placed in the backrest upholstery (pic 45-46). You can adjust the position of the belt by putting the straps into the different holes (higher or lower ones). Fasten the straps behind the backrest by use of Velcro (pic. 47-48).





AKCESMED

STEP 2. Pass the side straps through the stroller frame (pic.49), and then slide them into the plastic strap adjuster (pic.50-51). Repeat on the other side of the stroller.



STEP 3. Afterwards, pass the remaining strap with buckle (1) as indicated in the pictures below (pic. 52-54).





🕾 +48 17 864 04 77 🛛 🖂 e





STEP 4. To fasten the trunk stabilizing belt, slide the strap (1) into the other strap buckle (2), and then fasten it by use of Velcro (pic. 55,56). Depending of the width of the trunk, you can use shorter (2) or longer strap (3) with buckle (pic 55).





6.4. **REF ULE_115** Calf belt

In order to mount the calf belt, you should:

STEP 1. Attach the brace to the footrest frame.



6.5. **REF** ULE_121 Head-neck support

This kind of headrest is highly indicated for patients with head-dropping problems. In order to mount the headrest, put the headrest straps into the holes placed in the backrest upholstery. Fasten the strap behind the backrest using Velcro.

STEP 2. Fasten the brace behind by use of Velcro.









6.6. **REF** ULE_124 Loins support

In order to mount the Loins support, you should:

STEP 1. Put the Loins support on the backrest and interlace the straps through the holes in the upholstery (pic. 60, 61).





STEP 2. Fasten the belts behind the backrest (pic. 62, 63).





6.7. REF ULE_126 Trunk belt

The trunk belt can be easily cut to any length. In order to do that, you should:

STEP 1. Unfasten the belt from the back (pic. 64).



🕾 +48 17 864 04 77

export@akces-med.com





STEP 2. Choose the appropriate length of the belt and cut it along the designated place (pic. 65, 66).





STEP 3. Fasten the ends of the belt using Velcro fastener (pic. 67). The belt should be mounted to the stroller around the backrest (pic. 68).







In order to mount the abduction block, you should:

STEP 1. Remove the 5-point belts attached to the standard stroller equipment.

STEP 2. Slide the abduction block bar into the handle placed in the seat upholstery (pic. 69).

STEP 3. Tighten the abduction block to the seat by use of Allen screw (pic. 70, 71).











6.9. **REF ULE_136** Thigh belt

In order to mount the thigh belt, you should:

STEP 1. Put the belt on the stroller and fasten it under the seat using Velcro.





6.10. REF ULE_134 Narrowing pads (set) 6 cm REF ULE_137 Narrowing pads (set) 10 cm

Cushions can only be attached to the seat along with the rigid side supports (low or high) that are trimmed with one-sided Velcro material (see page 30, 31).









6.11. **REF ULE_148** High rigid trunk support

In order to mount the supports, you should:

STEP 1. Slide the support between the frame and the seat. Then, press the upper hook of the support down (pic. 76).

STEP 2. Press the other hook in the designated place and secure with Velcro strap (pic. 77, 78).









The supports are covered with one-sided Velcro material which means that you can attach to it any patient-stabilizing elements that are equipped with Velcro e.g. seat narrowing cushions.

29





6.12. **REF ULE_153** Low latteral support

REF ULE_149Low latteral support with seat narrowing cushion 6cmREF ULE_150Low latteral support with seat narrowing cushion 10
cm

In order to mount the supports, you should:

STEP 1. Slide the support between the frame and the seat (pic. 80).



STEP 2 Then, mount the support using Velcro strap and hooks that need to be pressed to the frame (pic. 81,82).





The supports are covered with onesided Velcro material which means that you can attach to it any patientstabilizing elements that are equipped with Velcro e.g. seat narrowing cushions.





6.13. REF ULE_114 H Vest REF ULE_130 4-points safety vest

In order to attach the vests to the stroller, you should:

STEP 1. Detach two bottom straps and the plastic strap adjusters (pic. 84). Then, interlace the remaining strap through the black holder placed next to the backrest (pic. 85).



STEP 2. Slide the end of the strap into the plastic strap adjuster located on the other side (pic. 86-88).













STEP 3. Slide the strap into the remaining plastic strap adjusters (the ones you previously detached in STEP 1, pic 89).

STEP 4. Slide the strap into the snap buckle in the bottom part of the vest (pic 84, 85). Afterwards, slide again the end of the strap into the strap adjusters (pic. 86, 87).











STEP 5. Attach the upper part of the vest at the desired height and secure it with the knob from the back (pic. 88, 89).







www.akces-med.com 🖀 +48 17 864 04 70 🖂 produkcja@akces-med.com





Properly attached H vests and 4 point vest.





6.14. **REF ULE_125** 6-point safety vest

To attach the vests, you should:

STEP 1. Follow the H vest and 4 point safety vest instruction (see page 32-33, STEP 1-5).

STEP 2. The 6-point safety vest has additional security in the form of a long strap with snap buckle that should be passed around the backrest and fastened from the other side of the vest (pic. 92-95).









Properly attached 6-point safety vest.



The 6-point safety vest has a wide range of adjustments that allows you to easily fit it to the patient (pic. 97-102).







6.15. REF ULE_402 Umbrella

In order to attach the umbrella to the stroller, you should:

STEP 1. Loosen the clamp and attach it to the stroller frame (pic. 103).

STEP 2. Tighten the knob (pic. 104).

STEP 3. Set the desired position of the umbrella using the button (pic. 105).





In order to mount the tray, you should:

STEP 1. Place the tray hooks onto the selected frame elements (pic. 107-108).





^{6.16.} REF ULE_403 Tray





STEP 2. Afterwards, put the tray handles onto the frame and press them (pic. 109, 110).



STEP 3. In order to adjust the tray depth, loose the knob and set the desired depth. Then, tighten firmly the knob.





6.17. REF ULE_405 Canopy

In order to mount the canopy, you should:

STEP 1. Insert the canopy into the handles (pic.112-114).










STEP 2. The canopy is equipped with the case which can be attached to the rear of the backrest by use of Velcro straps (pic. 115, 116).





6.18. REF ULE_411 Cup holder

In order to attach the cup holder to the stroller, you should:

STEP 1. Loosen the clamp and put it on the stroller frame (pic. 117).

STEP 2. Tighten the knob (pic. 118).

STEP 3. Set the desired position using the button (pic. 119).





fig.128





6.19. REF ULE_412 Elastico cushion seat REF ULE_419 Seat cushion (thighs shape) REF ULE_421 Seat cushion (pommel shape)

In order to attach the cushion to the seat, you should:

STEP 1. Put the cushion on the stroller and fasten the straps under the seat using Velcro (pic. 121, 122).





6.20. **REF ULE_432** Tilt angle adjustment of the footrest

In order to mount the conversion kit which allows tilt angle adjustment, you should:

STEP 1. Prepare necessary tools and components:

- Conversion kit
- Allen wrench 3 mm
- Allen wrench 6 mm







6.22. REF ULE_505 Under seat basket

Under seat basket can be mounted to the chassis frame by use of latches (pic. 132-135).















STEP 2. Install the conversion kit in the footplate holes (pic. 124, 125).





STEP 3. Set the tilt angle of the footrest while tightening M8 bolt (pic. 126, 127).





You can attach the bags to the stroller's backrest or push handle. It also can be worn on the shoulder thanks to its long, adjustable strap.















THE USE OF STROLLER



CAUTION! ULISES EVO[™] stroller has to be used with compliance to guidelines included in this user manual. Remember to use the product only with someone who is trained and knowledgeable. Do not forget to prepare both product as well as patient before using ULISES EVO[™] stroller!

7.1. Preparation of the patient

Before using the product, it is essential to make sure that:

- · patient's outfit is light and comfortable sporty clothes are advisable;
- \cdot shoes should be comfortable and, moreover, they should stabilize ankle;
- due to hygienic reason and to avoid abrasion it is advisable for the patient to always wear clothes when using the device, skin must not be naked especially in places of regular contact with upholstery;
- drains, bags, catheters etc. if applied, have to be secured to function properly and not to disturb during using the product;
- if it is necessary for a patient to wear orthopedic supply (collar, corset, orthesis, shoes) if it is recommended by doctor.



CAUTION! Outfit should protect patient's skin from direct contact with upholstery as well as other parts of equipment.

CAUTION! Only doctor can make a decision concerning using any kind of orthopedic supply!





7.2. How to use stroller

- 1. Before use it is necessary to:
 - Unfold the device;
 - · Place the device on even ground;
 - · Make sure that all parts are serviceable, correctly mounted and stable;
 - Block brakes;
 - · Adjust the device to patient's measurements;
 - Undo vest, trunk belts etc so as to place easily the patient is the device;
 - Place patient in the device;
 - Fasten all elements which stabilize patient e.g. vest, trunk belts, feet stabilizing belts etc.;
 - Mount additional equipment e.g. tray.
- 2. In case of patients who have problem with bladder, it is necessary to put special cloth on seat. This cloth can be purchased in pharmacy or medical shops.
- 3. It is essential to hold stroller with the aid of handle during using it otherwise stroller can roll.
- 4. Remember to place stroller forwards
- 5. In case of any obstacles such as step or curb, it is necessary to raise front wheels by pressing the handle down . Next, place front wheels on step and then raise back part of stroller with the aid of handle.



CAUTION! Before the product is used, it is highly necessary to make sure that all belts as well as safety equipment are correctly fastened as well as adjusted. What is more, check if they do not become undone due to the pressure and, moreover, make sure that clampings from adjustable elements are firmly tightened!

Equipment works correctly only when the device is adjusted according to patient's height as well as weight. Remember to take all these measurements into account while purchasing the product. Correct adjustment is the most important due to the fact that only then the product is safe. It is forbidden to force child to keep specific position because he/she will feel uncomfortable and it may lead to wrong posture.





8 SAFETY RULES



CAUTION! ULISESevo™ stroller has to be used in compliance with guidelines in this user manual.

It is highly important to consult either doctor or physiotherapist in order to make right decision concerning the choice of stroller and its size. It is also essential to familiarize with the guidelines described in user manual before using the device.

- 1. Before use it is necessary to:
 - make sure that there are no damages or different faults which could endanger patient's life; check if all parts are working and are correctly mounted and, moreover, check if they are not broken;
 - check if all screws, knobs and all glued, sewed or tapped parts are tightened;
 - · check if brakes are operating correctly;
 - · check if buckles are operating correctly;
 - · always use safety belts, Velcro tapes should overlap min. 8cm;
 - adjust the device according to patient's measurements and needs e.g. the adjustment of pelottes, footplate etc.;
 - make sure that all movable parts and knobs for adjustments are stable and locked;
 - secure patient with the aid of elements such as vest, foot stabilizing belts, pelottes, abduction block, headrest. It is highly important to secure patient because it safeguards him/her from falling out. If abduction contracture occurs, it is recommended to use abduction block in order to prevent ulcers between knees.





- 2. It is forbidden to use stroller for different purposes that it is intended to.
- 3. It is forbidden to use stroller when it is damaged or any part is missing.
- 4. It is forbidden to leave the patient unattended.
- 5. Stroller should be used only when there is someone who is trained and knowledgeable.
- 6. It is forbidden to leave stroller and the patient on unstable ground unattended, even if the brake is engaged.
- 7. It is forbidden to either go up or down the stairs; it doesn't matter if the patient is placed in the device or not.
- 8. It is important to engage brake during getting into and getting out the stroller.
- 9. Make sure that all limbs are not close to spokes or other adjustable elements.
- 10. Although the device and the upholstery are made of fire-resistant materials, please pay your attention when approaching to sources of fire or sources of high temperature during using the device. Moreover, wheelchair shouldn't be stored in rooms with high humidity.
- 11. Secure stroller during transportation in order to avoid any possible damages. Manufacturer doesn't take responsibility for damages which appeared due to improper preparation of the device during transportation.





CAUTION! It is forbidden to place the patient in the device while being in car, bus or train.



CAUTION! The product contains small parts which children may swallow.

CAUTION! It is forbidden to make any changes concerning stroller 's structure due to the fact that it is dangerous and it results in loosing warranty.

CAUTION! The weight of user should not be bigger than maximal weight described on page 12 where you will find information regarding this issue.

- 1) AKCES-MED Sp. z. o. o sells products free from manufacturing defects or faults.
- 2) Manufacturer gives the purchaser 24-month guarantee from the date of purchase. All defects revealed during warranty period will be repaired free of charge.
- 3) All defects which appear after warranty period can be repaired only when the user will pay for it. This rule doesn't apply to refunded products user can ask National Health Refund for refund.
- 4) Manufacturer claims that the product can be used for 48 months. User can use this product after this period only when he/she will send the device to manufacturer so as to service it. Next, manufacturer determines again how long the product can be used and, moreover, the date of the next service will be determined.
- 5) If product is damaged, please contact our dealer or AKCES-MED customer service (ph. 17/8640470)







MAINTENANCE AND CLEANING

Stroller should be kept clean and used in compliance with the manufacturer's recommendations. The device should be kept in clean, dry and ventilated room. Maintenance as well as cleaning have to be performed at least once a month. Regular cleaning and maintenance of stroller are indispensable to maintain its correct functionality as well as safety.

It is highly advisable to remember about proper maintenance in order to extend product's aesthetics. So as to do that, it is essential to comply with these guidelines:

- It is recommended to clean upholstery with the aid of cleaning product, soft sponge and warm water. If you use sponge which is too wet, stain may get bigger, upholstery may be destroyed. It is necessary to remember that the excess of dampness should be removed from fabric before next cleaning. Please rinse upholstery with warm water and clean white cloth once stain is removed.
- 2. It is forbidden to soak upholstery!
- 3. It is forbidden to clean upholstery with the aid of hard brushes.
- 4. It is forbidden to clean upholstery with strong chemical agents such as solvent or caustic chemicals.



CAUTION! Akces-MED. Sp z o. o takes no responsibility for damages caused by use of wrong clearing agent.

- 5. It is forbidden to soak upholstery during cleaning. Remember that upholstery has to be always dry before using.
- 6. It is forbidden to clean, make any changes and maintain wheelchair during using it.
- 7. All metal as well as plastic elements should be cleaned by hand with the aid of general cleaning products. Remember to familiarize yourself with cleaning product's leaflet.







CAUTION! Remember that upholstery has to be always dry before using!

Before using it is necessary to:

- clean wheels;
- check if base as well as wheels are correctly mounted; check if brake stops turning wheel;
- make sure that all wooden parts e.g. tray don't have splinters;
- check if wheelchair construction is free of breaks or any kinds of deformities;
- check if all movable parts are working and all screws, nuts, knobs etc. are correctly tightened;
- check if belts, vests, trunk belts, feet stabilizing belts etc are not frayed; check if buckles and clamps work correctly.



CAUTION! Manufacturer takes responsibility only for hidden faults or damages which appeared despite the fact that user complied with the guidelines.

CAUTION! In case of any damages, it is essential to contact the person who sold you the product.





ORTHOPAEDIC DEVICES MANUFACTURER



Akces-MED Sp. z o.o. Jasionka 955B 36-002 Jasionka tel. (17) 864 04 77

www.akces-med.com e-mail: export@akces-med.com

Authorised Dealer



Date of issue: 02.02.2015 The last update: 31.08.2017 Dur passion – your independence!